

8 December 2010

Dear Health Care Professional,

**Eli Lilly Australia supplied BD Micro-Fine™+ 0.25mm (31G) x 8mm Pen Needles
(Item Ref 320196, Lot 1226252 and Item Ref 320510, Lot 5005173)**

It has come to the attention of Eli Lilly Australia Pty Ltd that quantities of BD Micro-Fine™+ 8mm Pen Needles (Item Ref 320196, Lot 1226252 and Item Ref 320510, Lot 5005173) have recently been supplied by Lilly sales representatives after their labelled expiry date. At this stage we understand that only Lot Numbers 1226252 and 5005173 are affected.

While we recognise that you would routinely examine the expiration dates on all pen needles, we wanted to specifically draw your attention to this matter.

We ask that you please check your stock for these specific lots of pen needle packs, isolate them and do not issue them to patients. Please return these needles to your Lilly Sales Representative or contact Lilly's Medical Information staff on (02) 9325 4622 to arrange for their collection.

Any medical inquiries, immediate concerns or questions should be directed to Lilly's Medical Information staff on (02) 9325 4622.

Yours sincerely

Eli Lilly Australia Pty Ltd



Dr. Deon Gouws

B.Sc; MBChB; Dip PEC; MBA; LLB

Medical Director

Eli Lilly Australia Pty Ltd